

Core Technology Co., Ltd.

Room 100, C Building, No. 29 Life Park Rd, Changping District, Beijing, 102206 China
Tel: 86-10-69390616 Fax: 86-10-69390660 Website: www.coretests.com

Certificate of analysis

Product Name	COVID-19 AgTest		
Lot No.	20201030	Specification	1pc/pouch, cassette
Test Basis	Product technical requirements	Detection Quantity	55PCS
Manufacturing Date	2020/10/30	Expiry Date	2022/10/29
Test Items	Standard Regulation		Test Results
Physical properties	The product and packaging should be clean, free from stains, the internal and external packaging logos should be consistent, the contents should be complete, and the structure should be complete.		Compliance
	The membrane strip width of the film strip should not be less than 2.5mm.		3.10 mm
	The liquid migration speed should not be less than 5mm / min.		14mm/min
Minimum detection amount	Test with the national reference material or the lowest detection limit sample S standardized by the national reference, after dilution with buffer or pure water by 1:50, 1:100, 1:200, 1:400, 1:800 and 1:1600 and mark them as S1-S6, S1-S4 are all positive, S5 and S6 can be positive or negative.		Compliance
Positive reference compliance rate	Test with the national reference material or 8 positive reference materials standardized by the national reference, after dilution with buffer or pure water by 1:10, all are positive.		Compliance
Negative reference compliance rate	Directly use the original multiple to test the national reference material or 20 negative reference materials standardized by the national reference material, all are negative.		Compliance
Repeatability	Test with the national reference material or repetitive sample R standardized by the national reference material, after dilution with buffer or pure water by 1:10, 1:100 and mark them as R1, R2. The above two concentrations are tested by 10 times respectively. The 10 test results of R1 and R2 should all be positive and the color rendering should be uniform.		Compliance
Test Conclusion: This product is tested according to the product technical requirements of COVID-19 Ag Test and the results are qualified.			
Inspected by: Lin Li Date: 2020/11/05	Reviewed by: Yuanyuan Wang Date: 2020/11/05	Approved by: Yuanyuan Wang Date: 2020/11/05	